

Certificate



Certificate no.

CU 72092765 01

License Holder:

Dragon Laboratory Instruments
Limited
F3, Block 12, Yu Hua Rd. 28, Area B
Konggang Ind.Development Zone

101318 Shunyi District, Beijing
China

Manufacturing Plant:

Dragon Laboratory Instruments
Limited
F3, Block 12, Yu Hua Rd. 28, Area B
Konggang Ind.Development Zone

101318 Shunyi District, Beijing
China

Test report no.: USA-DS 30983640 001

Client Reference: Lei Zhang

Tested to: UL 61010-1:2004
CAN/CSA-C22.2 NO. 61010-1-04

Certified Product: Magnetic Stirrer

License Fee - Units

Model Designation: 1) MS-H Pro, MS-H-S
2) MS-Pro
3) MS-S

7

Rated Voltage: AC 100-120V, 50/60Hz
Rated Power: 1) 550W
2) 50W
3) 30W

Protection Class: I

7

Appendix: 1, 1-3

Licensed Test mark:



Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue
(day/mo/yr)

07/01/2010